



書類NO. 169-01

KEY NO.	ブヒンメイ	PART NAME	販売区分 CATEGORY	CODE NO. コードNO.						備考 COMMENTS
				169-205 PDM-75	169-207 PDM-100	169-203 PDM-1"	169-204 PDM-2"	169-206 PDM-3"	169-208 PDM-4"	
1	アンビル	ANVIL	Y	201363						
2	アンビル(スピンドルカワ)	ANVIL(SPINDLE SIDE)	Y	201294						
3	ネジ	SCREW	Y	300684						
4	キーヒス	KEY SCREW	Y	201194						注記/Note
5	スリワリツキトメネジ	SETSCREW		201193						
6	テーパナット	NUT,ADJ	Y	04GAA115						
7	カイトウ	SLEEVE	Y	200216	200217	200116				
8	スピンドルクミタテ	SPINDLE ASSY	Y	980699		980700				
8-1	スピンドル	SPINDLE	Y	303788						
8-2	スピンドルネジブ	THREADS,SPINDLE	Y	303774		303775				
8-3	ウケコマ	SETTING PIECE	Y	303776						
9	シンブル	THIMBLE	Y	303791		303792				
10	シンブルキャップ	CAP,THIMBLE		303777						
11	ラチェット	RATCHET STOP		04GZA240						
12	ホーネツカバークミタテ	COVER ASSY		-----		951639	951640	-----		
12-1	サラコネジ(M2.6×8)	SCREW(M2.6×8)		-----		306218		-----		
13	ネームプレート	NAMEPLATE		951345	951346	-----		951347	951348	
14	スパナ	SPANNER		301336						
	トリアツカイセツメイシヨ	USERS MANUAL		99MAA019M						注記/Note
	【包装資材】	【PACKING】								
	カクノウハコ	HOUSING CASE		475116	475117	475114	475115	475116	475117	
	ケシヨウケース	PAPER CASE		471431	471432	471685	471430	471431	471432	
				KeyNo.4: 嵌合、精度に影響がでる恐れあり修理対応が望ましい						
				KeyNo.4: The replacement of the key screw may impair the movement of the spindle and the accuracy of the micrometer. Therefore, it is recommended to ask for repair.						
				取扱説明書: 記載追加						
				Users Manual : The users manual has been added to this parts list.						

169

OUTSIDE MICROMETERS
Series 169
DISC TYPE, NON-ROTATING SPINDLE

Ref. No. 169-01B-2016-FEB

169シリーズ(PDM)
直進式歯厚マイクロメータ

169